

µC Based Blood Refrigerator Controller



- WTC-2201-BLSize: 96(w)x96(h)x65(d) mm
- WTC-2001-BLSize: 96(w)x48(h)x110(d) mm
- WTC-2101-BLSize: 72(w)x72(h)x65(d) mm
- WTC-2301-BLSize: 48(w)x48(h)x85(d) mm

- Microcontroller based,
- Four digit, dual display
- Supply 100-260 VAC
- Input: RTD PT100
- Output: 1. Relay Contact or SSR, 2. Alarm contact or Buzzer
- On/Off Control action
- Cooling Application
- Range: -99.9 to 100.0°C
 - Accuracy of 0.5% of full scale
- Temperature As per range adjustable in set mode.
- Alarm High 0.3 to 10.0 adjustable
- Alarm Low 0.3 to 10.0 adjustable
- Alarm mute by Decrement key
- Hysteresis 0.2 to 25.0°C adjustable
- Delay 0 to 100 minute adjustable for compressor
- OPEn display if Sensor break
- Relay trip if sensor breaks, connection open
- Calibration Error correction by front key switch (No preset setting)
- Configure link @ Back terminal to avoid unauthentic entry.
- Defrost Cycle start temperature adjustable(optional)
- Defrost On time 0 to 250 Minute
- Defrost Off time 1 to 9999 Minute
- Defrost on indication by dp blinking
- Defrost on time countdown indication (Double display)
- Manual defrost by pressing increment key for 5 second

WAVE INSTRUMENTS LLP

L-45/9, Phase-3, Opp. Mazda Ltd., GIDC Naroda, Ahmedabad -382330, Gujarat – INDIA

☎ +91- 9722003425, ☎ +91-79-40043425, 📞 +91-7043003425

Email: waveinstruments@gmail.com, Website: <http://www.wavetronics.co.in>